



SGS Canada Inc.
 657 Consortium Court
 London - Ontario - N6E 2S8
 Phone: 519-672-4500 FAX: 519-672-0361

27-July-2020

Mun of West Elgin (Recreational Water)

Attn : Adam Ecker

Date Rec. : 21 July 2020
LR Report: CA30310-JUL20

22413 Hoskins Line, PO Box 490,, Rodney
 , NOL 2C0
 Phone: 519-319-1033, Fax:

Copy: #1

CERTIFICATE OF ANALYSIS


Final Report

Sample ID	Sample Date & Time	Temperature Upon Receipt °C	E.Coli cfu/100mL	Fecal Streptococcus cfu/100mL
1: Analysis Start Date		---	21-Jul-20	21-Jul-20
2: Analysis Start Time		---	13:35	13:35
3: Analysis Completed Date		---	24-Jul-20	24-Jul-20
4: Analysis Completed Time		---	10:41	10:41
5: PWQO Limits		---	100	200
6: Area #1	21-Jul-20 08:30	19.4	6	122
7: Area #2	21-Jul-20 08:30	19.4	8	94
8: Area #3	21-Jul-20 08:30	19.4	14	102
9: Area #4	21-Jul-20 08:30	19.4	14	112
10: Area #5	21-Jul-20 08:30	19.4	18	104

PWQO/ IBFR- Provincial Water Quality Objectives and/or International Blue Flag Requirements

Method Descriptions

Units	Description	SGS Method Code
cfu/100mL	E.Coli by MF (using mFC Basal-BCIG)	ME-CA-[ENV]MIC-LAK-AN-006
cfu/100mL	Fecal Streptococcus by Membrane Filtratio	ME-CA-[ENV]MIC-LAK-AN-008


Cristal Schuster
 Project Specialist-London,
 Environment, Health & Safety